

Assay Reagent Kit	Xpert PML-RARA	
Catalog Number	RUO: RPMLRARA-10	
Technology	Nested RT-qPCR	
Targets	PML-RARA	
Batch or On-Demand	On-demand	
Minimum Batch Size	1	
Sample Type	Peripheral blood (EDTA)	
Sample Volume	4 ml	
Sample Extraction	Automated/ integrated	
Precision Pipetting	Not Required	
Off-board Sample Preparation Time	Approximately 30 minutes	
TAT	Approximately 3 hours (including 30-minute off-board sample preparation time)	
Internal Controls	Endogenous control (ABL1) ✓	Probe Check Control (PCC) ✓
Limit of Detection/ Sensitivity (EDTA)	0.012%	
Linear Range	0.012% – 500% PML-RARA/ABL	
System & Software	GeneXpert Dx System GeneXpert Dx software version 6.4 or higher	
Sample Stability	2–8 °C for up to 72 hours	
Kit Storage	2–8 °C	
Commercial Controls	Refer to Instructions for Use (IFU) or Contact Cepheid Technical Support	

For Research Use Only. Not for use in diagnostic procedures.

CORPORATE HEADQUARTERS

904 Caribbean Drive
Sunnyvale, CA 94089 USA

TOLL FREE +1.888.336.2743
PHONE +1.408.541.4191
FAX +1.408.541.4192

EUROPEAN HEADQUARTERS

Vira Solelh
81470 Maurens-Scopont France

PHONE +33.563.82.53.00
FAX +33.563.82.53.01
EMAIL cepheid@cepheideurope.fr

www.Cepheid.com

© 2024 Cepheid. 10113-01